



Enquiry No. : CHE/DK-GD/SERB/2013/03

December 12, 2013

Quotations are invited from interested suppliers for: **Carbon Monoxide Monitor and Alarm**

The technical and commercial bids are to be provided in separate envelopes. The Carbon monoxide monitor and alarm is for laboratory use in a continuous flow system.

Specifications for Carbon Monoxide Monitor and Alarm:

Range: 0-1000 ppm (adjustable in 100 ppm
increments)

Display: Digital

Repeatability: 2%

Power: 220-230 V AC

Resolution: 1 ppm

Accuracy: $\pm 2\%$ full scale

The carbon monoxide monitor should have the ability to give audio alarm when CO concentration falls below a user preset value.

Kindly mention about the terms and conditions such as;

- * Validity of the quotation for a minimum of 60 days.
- * Warranty period
- * Mention the FOB rate (if imported).
- * Special educational discounts, if any.

For indigenous or import items please use the form number 603 or 608 respectively given at

<http://web.iitk.ac.in/dord/rndforms/603.pdf>

<http://web.iitk.ac.in/dord/rndforms/608.pdf>

The quotations are to be sent to the under signed by December 23, 2013. The envelope should mention
“**QUOTATIONS FOR CO Monitor and Alarm**”

Goutam Deo

Professor, Department of Chemical Engineering

IIT Kanpur, Kanpur- 208 016