

To,

Ref No: IIT/SIIC/Mole/TMCCEC/6/24.11.14

**CORRIGENDUM NOTICE**

The descriptions of following items in the BOQ of above quotation may please be read as following:-

S.No.	Item No.	Mentioned in the quotation	May please be read as
1	Qty. 2	Milli Ohm and Micro Ohm Meter (Upto 5A)	<b>Milli Ohm and Micro Ohm Meter</b> <ul style="list-style-type: none"> <li>• Basic accuracy 0.25%</li> <li>• Max. Test Current : 5A (120mΩ)</li> <li>• Manual or Auto Range</li> <li>• Four Terminal Kelvin Measurement</li> <li>• 6 Ranges with 1μΩ best resolution</li> <li>• Memory of 3000 Measurements data</li> <li>• Large LCD with Backlight &amp; HOLD Function</li> </ul>
	Qty. 2	Milli Ohm and Micro Ohm Meter (Upto 10A)	<b>Milli Ohm and Micro Ohm Meter</b> <ul style="list-style-type: none"> <li>• Basic accuracy 0.25%</li> <li>• Max. Test Current : 10 A (60mΩ)</li> <li>• Manual or Auto Range</li> <li>• Four Terminal Kelvin Measurement</li> <li>• 6 Ranges with 1μΩ best resolution</li> <li>• Memory of 3000 Measurements data</li> <li>• Large LCD with Backlight &amp; HOLD Function</li> </ul>
2	Qty. 3	Clamp-On leakage Current Tester	<b>Clamp-On leakage Current Tester</b> <ul style="list-style-type: none"> <li>• 1μA Ultra High Resolution</li> <li>• Large Conductor Diameter (30mm)</li> <li>• AC current in 5 Ranges (40mA, 400mA, 4A, 40A, 60A)</li> <li>• Wide frequency Response (40 Hz ~ 1KHz)</li> <li>• AC A, AC V, Ω and continuity</li> <li>• Large 3 ¾ Digit LCD (4000 count)</li> <li>• 600 Vrms Overload protection in Resistance Range</li> </ul>
	Qty. 3	Clamp-On Earth/round Resistance & leakage Current	<b>Clamp-On Earth/round Resistance &amp; leakage Current Tester with Jaw Opening 23 mm.</b>

		Tester with Jaw Opening 23 mm	<ul style="list-style-type: none"> <li>• Non Contact Ground Resistance Measurement</li> <li>• Ground Resistance Measurement 0.01Ω - 1000Ω</li> <li>• Suitable for Grounding with the flat strip (Jaw sized 65 mm)</li> <li>• Leakage current (0.5mA ~ 30.00A)</li> <li>• Auto Ranging</li> <li>• Continuity Loop test</li> <li>• Auto Power Off</li> <li>• Jaw Size :- 65x 32 mm Approx.</li> </ul>
3	Qty.4	Flexible AC current probes – 3000A AC	Flexible AC current probes – 3000A AC <ul style="list-style-type: none"> <li>• Range- 3000 a AC</li> <li>• Output :- 100mV/1KA AC</li> <li>• Accuracy :- ± 1%</li> <li>• Probe Length :- 600 mm</li> <li>• Output Terminal :- 4 mm Banana pin</li> </ul>
	Qty. 4	Clamp on CT suitable for power & Harmonics Analyzer (Model 5850) of range : 1/10/100 A Ac one set= 3 Pcs (R,Y,B)	Clamp on CT suitable for power & Harmonics Analyzer of range : 1/10/100 A Ac one set= 3 Pcs (R,Y,B)
4	Qty. 7	3 Phase calibrator of Sinusoidal Currents & Voltages	3 Phase calibrator of Sinusoidal Currents & Voltages. <ul style="list-style-type: none"> <li>• Stability DC : &lt; 0.01 % FS /2 Min, AC : &lt; 0.03 %FS/2 Min</li> <li>• Ripple : &lt; 0.1 %</li> <li>• Power Supply : 230 VAC, 50(60) Hz</li> <li>• Power Consumption : &lt; 180 VA</li> <li>• Relative Humidity : ≤ 80 %</li> </ul>

**Terms & Condition**

**1. The scope includes:**

- a. Setting up machine in IIT Kanpur.
- b. Initial Installation and configuration.
- c. Training.



---

2. **Bid Price**

- a) The contract shall be for the full quantity as described above. Corrections, if any, shall be made by crossing out, initialing, dating and re writing.
- b) All duties, taxes and other levies payable on the raw materials and components shall be included in the total price. **Except Central Excise Duty & CDEC** (custom duty), as IIT Kanpur is exempted from this duty.
- c) Sales tax in connection with the sale shall be shown separately.
- d) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- e) The Prices shall be quoted in Indian Rupees only.

3. Each bidder shall submit only one quotation.

4. **Validity of Quotation**

Quotation shall remain valid for a period not less than 60 days after the deadline date specified for submission.

5. **Evaluation of Quotations**

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which

- (a) are properly signed ; and
- (b) Conform to the terms and conditions, and specification

**The Quotations would be evaluated separately for each item**

Sales tax in connection with sale of goods shall not be taken into account in evaluation.

6. **Award of contract**

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive (includes technically suitable) and who has offered the lowest evaluated quotation price.





**SIDBI INNOVATION AND INCUBATION CENTRE  
INDIAN INSTITUTE OF TECHNOLOGY KANPUR**



**Gross Total Cost : Rs. ....**

We agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs. ....(amount in figures ) (Rs. .... amount in words) within the period specified in the Invitation for Quotations.

We also confirm that the normal commercial warrantee/guarantee of 60 months shall apply to the offered goods.

**Signature of Supplier**

**SPECIAL CONDITION**

**1) Authorization from Manufacturer**

In the case of a Bidder offering to supply goods under the contract which the Bidder did not manufacture or otherwise produce, the Bidder has been duly authorized by the goods' Manufacturer or producer to supply the goods in India.

**2) Proof of Manufacturing and past performance.**

Details of experience and past performance of the bidder on equipment offered and on those of similar nature within the past one year and details of current contracts in hand and other commitments

**Dr. B.V Phani**

Associate Dean, Innovation & Incubation

Associate Professor, Finance, Innovation & Entrepreneurship, IME

Coordinator, SIDBI Innovation & Incubation Centre

Coordinator, Syndicate Bank Entrepreneurship Research and Training Centre

Email: [bvphani@iitk.ac.in](mailto:bvphani@iitk.ac.in)/[siic@iitk.ac.in](mailto:siic@iitk.ac.in)

Tel. No.:- 0512-259-7979

Fax No.:- 0512-259-7553

**SIDBI INNOVATION AND INCUBATION CENTRE, INDIAN INSTITUTE OF TECHNOLOGY KANPUR, KANPUR-208016**

Phones: 91-512-2597057/2597979(O) Fax: 91-512-2596177

e-mail: [bvphani@iitk.ac.in](mailto:bvphani@iitk.ac.in) web: [www.iitk.ac.in/siic](http://www.iitk.ac.in/siic) <http://home.iitk.ac.in/~bvphani/>