

ENQUIRY NO.IITK/CHM/JNM/12-13/27 DATED: 11/02/2013

The suppliers are requested to submit quotations for

CYCLIC VOLTAMMETRY EQUIPMENT

by 20th February, 2013 (Wednesday) to the undersigned in a sealed envelope.

The price quoted should be FOB or INR.

Potentiostat/Galvanostat with the following specifications.

1. Capable of performing the following experiments: Cyclic voltammetry, Linear sweep voltammetry, Normal pulse voltammetry, Differential pulse voltammetry.
2. 1 no. of channels and compliance voltage of ± 12 V.
3. Maximum current of 100 mA; applied potential of ± 3.275 V with 100 μ V resolution.
4. Applied current: 1–50 mA Range: ± 10 V.
5. Tool kit and associated softwares.
6. Cell stand with magnetic stirrer.

Dr. J. N. Moorthy
Department of Chemistry
Indian Institute of Technology Kanpur
Kanpur 208016, INDIA